

**IEC****IECEE**

®

TM

Ref. Certif. No.

**FR\_711833**

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

**CB TEST CERTIFICATE**

Product

**Incubator**

real-time incubator and colony counting station

Name and address of the applicant

**INTERLAB**30 Chemin du bois des arpens  
78860 ST NOM LA BRETECHE  
France

Name and address of the manufacturer

**INTERLAB**Route de Jalenques  
15340 PUYCAPEL  
France

Name and address of the factory

**INTERLAB**Route de Jalenques  
15340 PUYCAPEL  
France

Note: When more than one factory, please report on page 2

 Additional Information on page 2

Ratings and principal characteristics

100-240V~ - 50/60Hz - 11.5A

Trademark / Brand (if any)

INTERSCIENCE

Customer's Testing Facility (CTF) Stage used

/

Model / Type Ref.

ScanStation® 100  
ScanStation® 200  
ScanStation® 300

Additional information (if necessary may also be reported on page 2)

/  
 Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 61010-1:2010 + A1:2016  
IEC 61010-2-010:2019  
IEC 61010-2-081:2019

As shown in the Test Report Ref. No. which forms part of this Certificate

158483-749591-M Version : 01  
158483-749591-N Version : 01  
158483-749591-O Version : 01

This CB Test Certificate is issued by the National Certification Body

LABORATOIRE CENTRAL DES INDUSTRIES ELECTRIQUES - LCIE  
33 avenue du Général Leclerc  
92260 Fontenay-aux-Roses, FRANCE  
[www.lcie.fr](http://www.lcie.fr)

Date: 30/06/2021

Signature: **Gilles LEMONNIER**  
Certification Officer